

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 02 2011
2011
3/9/11

DataMaster cdm S/N 130195

Supervisor/Operator Performing the Verification Procedure:

A Name Robert J. Ely ID# 2681 ✓ Date 01-22-2011 ✓
Agency Craig Police Department Phone # (907) 826-3330
Instrument Location 506 2nd Street, Craig Alaska 99921

B Alco S/N X173355 ✓ Target Value .078 ✓ High Pressure 200 PSI
Alco Test Value Average .079 ✓ .078 ✓
1st Alco 2nd Alco

Signature Robert J. Ely

(OVER)

(Do Not write in the area below)

BMB
2/8/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows;

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons Responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR, 2011

Daniel L. Nicholas
(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

JANUARY 22, 2011

OPERATOR'S NAME:

ELV, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VER OF CAL

S.L. #: 123456789

DEPT/AGENCY: CR31

CASE/REPORT: 11-985450

TEST TYPE: V

ALCO TARGET VALUE: .078

ALCO S/N: X173355

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.88 in		
ALCO TARGET	.077	11:59
BLANK TEST	.000	11:00
INTERNAL STANDARD	VERIFIED	11:00
ALCO TO 29.88 in	.079	11:00
BLANK TEST	.000	11:01
SUBJECT SAMPLE	.000	11:01
BLANK TEST	.000	11:02
ALCO TO 29.88 in	.078	11:02
BLANK TEST	.000	11:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

JANUARY 22, 2011
TIME 18:56

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
SAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECT/PA	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' () * + , - . : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z ! " # \$ %